



广 微 测  
Gmicro Testing

## GUANGDONG DETECTION CENTER OF MICROBIOLOGY

### REPORT FOR ANALYSIS

Report №.

2020FM04645R01E

Name of Sample

Ozone Generator

Applicant

Highend Home Appliance Co.,ltd

Test Type

Entrustment Test



Address: Building 66, No.100 Central Xian Lie Road, Guangzhou, China

Postcode: 510070

Tel: +86 20 87137666

Fax : +86 20 87137668

Website : [www.gddcm.com](http://www.gddcm.com)



广微测  
Gmicro Testing

# GUANGDONG DETECTION CENTER OF MICROBIOLOGY

## REPORT FOR ANALYSIS



Report No.: 2020FM04645R01E Verification Code: 14978250

Name of Sample	Ozone Generator	Test Type	Entrustment Test
Applicant	Highend Home Appliance Co.,ltd	Address	Floor 3,16th-1,Meiyuan Road, Qianlong,Sanxiang town, Zhongshan City, Guangdong,China
Sample Source	Submitted for Testing by the Applicant	Sample Quantity	One Sample Submitted
Spec and Lot No of Sample	HE-151N	State and Characteristic	Machine
Sample Received Date	2020-03-09	Test Completion Date	2020-04-09
Test Standard and Method	Refer to GB 21551.3-2010 A		
Item Tested	Air virus elimination effect		
Test Conclusion	<p>The test data of the sample(s) is attached to the page(s) of this report.</p> <p>Issue Date: # DISTRIBUTED DATE# (Official Seal)</p>		
Remarks	<p>检验检测专用章</p>		

Editor: Wuqiuyan.

Verifier: Sun Jiusi

Approver: Xie Xiaobao





GUANGDONG DETECTION CENTER OF MICROBIOLOGY

ANALYSIS AND TEST RESULT

Report No.: 2020FM04645R01E

Virus	Action Time	Serial Number	Air virus content (TCID <sub>50</sub> /m <sup>3</sup> )	Removal rate (%)
H1N1 Influenza virus (A/PR/8/34) Host cell: MDCK	0 (CK)	1	1.80×10 <sup>6</sup>	
		2	3.77×10 <sup>6</sup>	
		3	2.86×10 <sup>6</sup>	
	1h	1	1.34×10 <sup>4</sup>	97.81
		2	1.80×10 <sup>4</sup>	98.45
		3	2.27×10 <sup>4</sup>	97.66

Note: The natural decay of the microorganisms in the air has been eliminated.

(Blank Below)

分析  
专用

Remarks	The machine was tested in a 30m <sup>3</sup> test chamber for 1 hour with the Ozone mode block on.
---------	--



## Notice Items

1. The Test report is invalid if not affixed with Authorized Stamp of Test and Paging Seal.
2. The Test report is invalid without signature of verifier and approver.
3. The Test report is invalid if being supplemented, deleted or altered.
4. Without prior written permission, the report cannot be reproduced, except in full.
5. Unless otherwise stated, the results shown in this test report refer only to the sample(s) submitted.
6. Any dispute of the report must be raised to the testing body within 15 days after the report is received, exceeding which the dispute will not be accepted.
7. For the tested sample(s) submitted by the applicant, the sample information in the test report is provided by the applicant and the laboratory is not responsible for its authenticity.
8. This test report is for reference only to the applicant and does not have a proof of effect for others.

